



EUREKA INDUSTRIAL EQUIPMENT PVT. LTD.

'J' 501, M. I. D. C., Pimpri, Pune - 411 018.

Cell : +91 9325751737 E-mail : works@eurekaflow.com Web : www.eurekaflow.com

SERVICE / COMMISSIONING / VISIT REPORT

REPORT NO.	DATE - 23-11-24	DATE OF VISIT : 23-11-2024
CLIENT : M/s. Shree Bhaivarnath Feeds Private Limited, Gate-no-704 Karde, Tal - Shirur Dist - Pune 412210		INVOICE NO :
VISITED PERSON NAME : Rohit Mapari		MODEL : Eumag - 25-T-S-II-CS-1-A15-230-M
DESIGNATION :		SR. NO. 2023-2024 / EMFM / 12515
TEL. : 9158256647		ATTENDED BY : Akshay Kadam

OBSERVATIONS Check this meter 2023-2024 / EMFM / 12515
 Check the calibration of this meter but it is not calibrating properly.
 Flow is on but customer does not know how much flow wants if it's calibration setting is disturbed.

SERVICE RENDERED :

Suggested to please send to meter our factory for ~~for~~ calibration purpose.

REMARKS
 Flowmeter is used for Molasses service. Possibility of deposition on electrodes. To be checked in factory

CLIENT'S REMARK :

Akshay 23-11-2024
 SIGNATURE OF SERVICE TECH./ENGINEER

CUSTOMER'S SIGNATURE & DATE
[Signature] 23/11/24